Access IV Pharmacy

CADD PRIZM PCA Flow Sheet

Patient Name: _____ Medication: _____

1. Check the label for correct medication and discard date
2. Do you have to change the battery? (see the Bag Change Instruction Sheet)
3. Remember to Reset the Reservoir Volume by pressing ,<Enter> on the Res Vol Screen (The pump has to be off)
4. Open all clamps and restart the pump by pressing <START/STOP> and then pressing <YES>

Change Tubing: ☐ Every Bag ☐ Weekly ☐ _____

Battery Change: ☐ Every Bag ☐ Every 3-4 days ☐ _____

Call if you have any questions (831 or 866) 384- 8080

Date	Res Vol (ml)	Reset Yes/No	UNITS mg or ml	Conc. mg/ml	Cont Rate ☐ mg/hr ☐ ml/hr	Dem Dose ☐ mg ☐ ml	Demand Dose min	Max Dose / hr	Dose Given/ Attempted	Reset Yes/No	mg Given	Reset	Lock Level	Batt Chng	Pain Level (1-10)	Signature
													1			